OMRON

IM-HEM-6181-E-EN-02-12/2018

2829387-3B

Automatic Wrist Blood Pressure Monitor

RS4 (HEM-6181-E) Instruction Manual



Table of Contents

- Introduction Important Safety Information
- 1. Know Your Monitor
- Preparation
- 3. Using Your Monitor 4. Using Memory Function
- 5. Other Settings
- 6. Error Messages and Troubleshooting
- Maintenance 8. Specifications
- 9. Limited Warranty
- 10. Guidance and Manufacturer's Declaration

Introduction

Thank you for purchasing the RS4 Automatic Wrist Blood Pressure Monitor. This blood pressure monitor is to be worn on your wrist. This blood pressure monitor uses the oscillometric method of blood pressure measurement. When the cuff inflates, this monitor senses the pressure pulsations of the artery underneath the cuff. The pulses are called oscillometric pulses. The electronic pressure sensor displays a digital reading of blood pressure.

Safety Instructions

This instruction manual provides you with important information about the RS4 Automatic Wrist Blood Pressure Monitor. To ensure the safe and proper use of this monitor, READ and UNDERSTAND all of the safety and operating instructions. If you do not understand these instructions or have any questions, contact your OMRON retail outlet or distributor before attempting to use this monitor. For specific information about your own blood pressure, consult with your physician.

Intended Use

This device is a digital monitor intended for use in measuring blood pressure and pulse rate in adult patient population with wrist circumference ranging from 13.5 cm to 21.5 cm. This monitor detects the appearance of irregular heartbeats during measurement and gives a warning signal with readings. It is mainly designed for general household use.

Receiving and Inspection

Remove this monitor from the packaging and inspect for damage. If this monitor is damaged, DO NOT USE and consult with your OMRON retail outlet or distributor.

Important Safety Information

Please read the Important Safety Information in this instruction manual before using this monitor.

Please follow this instruction manual thoroughly for your safety. Please keep for future reference. For specific information about your own blood pressure, CONSULT WITH YOUR PHYSICIAN.

Indicates a potentially hazardous situation AWarning which, if not avoided, could result in death or serious injury.

• DO NOT use this monitor on infants, toddlers, children or persons who cannot express themselves.

 DO NOT adjust medication based on readings from this blood pressure monitor. Take medication as prescribed by your physician. ONLY a physician is qualified to diagnose and treat high blood pressure.

• DO NOT use this monitor on an injured wrist or a wrist under medical treatment.

 DO NOT apply the wrist cuff on your wrist while on an intravenous drip or blood transfusion

 DO NOT use this monitor in areas containing high frequency (HF) surgical equipment, magnetic resonance imaging (MRI) equipment, computerized tomography (CT) scanners. This may result in incorrect operation of the monitor and/or cause

an inaccurate reading. • DO NOT use this monitor in oxygen rich environments or near flammable gas.

 Consult with your physician before using this monitor if you have common arrhythmias such as atrial or ventricular premature beats or atrial fibrillation; arterial sclerosis; poor perfusion; diabetes; pregnancy; pre-eclampsia or renal disease. NOTE that any of these conditions in addition to patient motion, trembling, or shivering may affect the measurement reading

• NEVER diagnose or treat yourself based on your readings. ALWAYS consult with your physician.

. This product contains small parts that may cause a choking hazard if swallowed by infants, toddlers or children. **Battery Handling and Usage**

· Keep batteries out of the reach of infants, toddlers or children.

Indicates a potentially hazardous situation which, if not avoided, may result in minor or Caution moderate injury to the user or patient, or cause

damage to the equipment or other property. Stop using this monitor and consult with your physician if you experience skin irritation or discomfort.

 Consult with your physician before using this monitor on a wrist where intravascular access or therapy, or an arteriovenous (A-V) shunt, is present because of temporary interference to blood flow which could result in injury. Consult with your physician before using this monitor if you

have had a mastectomy. Consult with your physician before using this monitor if you have severe blood flow problems or blood disorders as cuff

inflation can cause bruising.

• DO NOT take measurements more often than necessary because bruising, due to blood flow interference, may occur. • ONLY inflate the wrist cuff when it is applied on your wrist. • Remove the wrist cuff if it does not start deflating during a

measurement. • DO NOT use this monitor for any purpose other than

measuring blood pressure

 During measurement, make sure that no mobile device or any other electrical device that emits electromagnetic fields is within 30 cm of this monitor. This may result in incorrect operation of the monitor and/or cause an inaccurate reading. • DO NOT disassemble or attempt to repair this monitor or

other components. This may cause an inaccurate reading. DO NOT use in a location where there is moisture or a risk of water splashing this monitor. This may damage this monitor. DO NOT use this monitor in a moving vehicle such as in a car. • DO NOT drop or subject this monitor to strong shocks or


vibrations • DO NOT use this monitor in places with high or low humidity or high or low temperatures. Refer to chapter 8.

- Ensure this monitor is not impairing blood circulation by observing the wrist while measurement is occurring. · DO NOT use this monitor in high-use environments such as
- medical clinics or physician offices. • DO NOT use this monitor with other medical electrical (ME) equipment simultaneously. This may result in incorrect operation of the monitor and/or cause an inaccurate reading. · Avoid bathing, drinking alcohol or caffeine, smoking,
- exercising and eating for at least 30 minutes before taking a measurement. Rest for at least 5 minutes before taking a measurement.
- · Remove tight-fitting, thick clothing and any accessories from your wrist while taking a measurement. • Remain still and DO NOT talk while taking a measurement.
- ONLY use this monitor on persons whose wrist circumference is within the specified range of the cuff.
- Ensure that this monitor has acclimated to room temperature before taking a measurement. Taking a measurement after an extreme temperature change could lead to an inaccurate reading. OMRON recommends waiting for approximately 2 hours for the monitor to warm up or cool down when the monitor is used in an environment within the temperature specified as operating conditions after it is stored either at the maximum or at the minimum storage temperature. For additional information of operating and storage/transport temperature, refer to chapter 8.
- DO NOT crease the wrist cuff excessively. • Read and follow the "Correct Disposal of This Product" in chapter 10 when disposing of the device and any used
- accessories or optional parts • DO NOT use this monitor after the durable period has ended.
- Refer to chapter 8
- Battery Handling and Usage DO NOT insert batteries with their polarities incorrectly
- aligned. • ONLY use 2 "AAA" alkaline batteries with this monitor. DO NOT use other types of batteries. DO NOT use new and used batteries together. DO NOT use different brands of batteries together.
- Remove batteries if this monitor will not be used for a long period of time.
- If battery fluid should get in your eyes, immediately rinse with plenty of clean water. Consult with your physician
- If battery fluid should get on your skin, wash your skin immediately with plenty of clean, lukewarm water. If irritation, injury or pain persists, consult with your physician.
- DO NOT use batteries after their expiration date.
- · Periodically check batteries to ensure they are in good working condition.
- ONLY use batteries specified for this monitor. Use of unsupported batteries may damage and/or may be hazardous to this monitor.

1. Know Your Monitor


1.1 Contents

Monitor, 2 "AAA" alkaline batteries, storage case, instruction manual



- a. Wrist cuff
- (Wrist circumference 13.5 cm to 21.5 cm)
- b. Display
- c. Battery compartment
- d. [Memorv] button e. [Morning average] button
- f. [START/STOP] button

1.3 Display and Symbols




Α	Date	and time o	lisplay		۷.	Preparation	
в		Memory s	symbol		2.1	Installing Bat	
D	Ø		nen viewing readings sto	ored in the memory.	1. Push down cover and		
C	-		pressure reading				
D			pressure reading				
_	Pulse	rate appears	Memory number after the measurement				
E	for app	proximately	e 🕞 button, the memory one second before the p		2.	Insert 2 "AAA" a indicated in the	
	the dis	play. Deflation	symbol				
F	-		ring cuff deflation.				
G	-	Heartbeat Flashes wh	t symbol ile taking a measureme	nt.	2	Class the bette	
н			average symbol nen viewing the weekly a	average readings	5.	Close the batte	
		taken in the	• •				
			thin 4 weeks appear on				
ı	메 AVG		value symbol nen viewing the average	readings of the	Not	9	
	latest 2 or		3 readings taken within 1		• When the " 🚛 " s you replace batter		
			ymbol (low) en batteries are low.		Tł	replace batteries, nen replace with 2	
J		Battery symbol (depleted)			da	s soon as batteries ate and time displa	
			nen batteries are deplete sion symbol		• Tł	eplacing batteries ne supplied batteri sposal of used ba	
	SYS A	Appears if	our systolic blood press or the diastolic blood pre			th local regulation	
κ		or above.			2.2	Setting the D	
	<u>کېد</u>		hypertension symbo nen the morning weekly			et your monitor to	
	DIA 📥	mmHg or a			measurement for tAs soon as batterie		
		Your monitor	ng indicator or has a built-in advance			splay of your moni	
	Ξ	that is used the correct	l as an aid in determinin height.	g if the monitor is at		Press and hold the	
L			symbol appears when yo ition relative to your hea			Press and hold through the year	
	Ξ	measureme	ent. When your monitor i ative to your heart, the h	is higher or lower	2.	Press the [STA	
	Ξ	bar will app	ear up above, or below			the month will f the month, day	
		guide correct wrist positioning. Irregular heartbeat symbol					
		Appears ale readings w	0	Normal Heartbeat		Year	
		irregular rh	ythm is Blood pressu	re		M	
		times durin measureme	0	Irregular Heartbeat			
М	\bigcirc	An irregular rhythm is d	efined as a	e A A A			
		rhythm that less or 25%	more	re			
		measuring	erage rhythm detected v blood pressure. If it cont	inues to appear, we		Press the [STA It automatically to	
		of your phy	d you to consult with and sician.	follow the directions	Not	9	
	合		It error symbol	your body is moving	ar	you need to reset nd hold the 👜 bu	
N		Appears along with readings when your body is moving during a measurement. If it appears, remove the wrist cuff, and wait for 2 to 3 minutes. Then, remain still and			ar 2.3	nd time as above s Blood Pressu	
		take another measurement.		To h	elp ensure an acc ress raises blood		
0	(OK)	Appears if	guide symbol (OK) the wrist cuff is wrapped	around the wrist	st	ressful times. easurements shou	
	0		prrectly while taking a measurement. It also appears hen viewing past readings.			is important to take king measuremen	
	Cuff wrap guide symbol (loose)				commended. emember to have a		
	\bigcirc	Appears if the wrist cuff is not wrapped around the wrist correctly while taking a measurement. It also appears when viewing past readings.			pr	adings for your ph ovide an accurate	
01	2 5 6 1		01 0	agament of	re	ease use the Bloo adings over a cert	
rte	erial hy	ypertensio	delines for the man n sion by office and home	-		e diary, visit www. Caution	
/eff	muuns	or hypertens	on by office and nome	biood pressure levels	• Av	/oid bathing, drink	
	Systoli	c Blood				nd eating for at lea est for at least 5 m	
	Pres	sure	≥ 140 mmHg	≥ 135 mmHg			
		ic Blood					

* European Society of Hypertension (ESH) and European Society of Cardiology (ESC).

A Warning

 NEVER diagnose or treat yourself based on your readings ALWAYS consult with your physician




symbol flashes on the display, it is recommended

- es, turn your monitor off and remove all batteries 2 new alkaline batteries at the same time. es are correctly installed, the year flashes on the play to set the date and time
- es will not delete previous readings. eries may have a shorter life than new batteries. patteries should be carried out in accordance ons.

Date and Time

- to the correct date and time before taking a
- the first time ies are correctly installed, the year flashes on the onitor
- or 👜 button once to adjust the year. l the 😰 button to quickly advance the years. l the 👜 button to quickly go backwards
- ART/STOP] button to confirm the year and Il flash. Repeat the same steps to adjust av hour and minute.



[ART/STOP] button to store the setting. turns off after 3 seconds

et the date and time, replace batteries or press button until the year flashes, and set the date

sure Measurement Tips

ccurate reading, follow these directions: od pressure. Avoid taking measurements during


- ould be taken in a quiet place. ake measurements at same times each day ents in the morning and in the evening is
- e a record of your blood pressure and pulse physician. A single measurement does not te indication of your true blood pressure. ood Pressure Diary to keep records of several ertain period of time. To download PDF files of w.omron-healthcare.com.
- nking alcohol or caffeine, smoking, exercising east 30 minutes before taking a measurement. minutes before taking a measurement.

2.4 Applying the Wrist Cuff

use for your measurements.

- Following steps are for applying the wrist cuff to your left wrist. When you take a measurement on your right wrist, apply the wrist
- cuff using the same steps as on your left wrist. • The blood pressure can differ between the right wrist and the left wrist, and the measured blood pressure values can be different. OMRON recommends to always use the same wrist for measurement. If the values between both wrists differ substantially, please check with your physician as to which wrist to

A Caution



- Apply the wrist cuff to your left Roll up sleeve. Make sure your sleeve is not rolled up too tightly on your arm. This may constrict the flow of blood in your arm.
- 2. Put your wrist through the cuff loop. Your palm and monitor display should both face upward.



- 3. Position the wrist cuff leaving a clearance of 1 to 2 cm between the cuff and the bottom of your palm
- 1 to 2 cm
- clothing Check to make sure that there is no gap between your wrist and the cuff.

your wrist. Do not apply over

Note

• Make sure the wrist cuff does not cover the protruding part of the wrist bone on the outside of your wrist.


2.5 Sitting Correctly

To take a measurement, you need to be relaxed and comfortably seated in a room with a comfortable temperature. Place your elbow on the table to support your arm.

- Sit in a chair with your legs uncrossed and your feet flat on the floor. • Sit with your back and arm
- being supported.
- Your monitor must be
- approximately the same level as your heart. If the monitor is

too high above your heart, your blood pressure will be artificially low. If the monitor is too low below your heart, your blood pressure will be artificially high.

• Relax your wrist and hand. Do not bend your wrist back, clench your fist, or bend your wrist forward



3. Using Your Monitor

3.1 Taking a Measurement

• To stop the measurement, press the [START/STOP] button once to deflate the wrist cuff.

A Caution


• DO NOT use this monitor with other medical electrical (ME) equipment simultaneously. This may result in incorrect operation of the monitor and/or cause an inaccurate reading. · Remain still and DO NOT talk while taking a measurement.

1. Press the [START/STOP] button. All symbols appear on the display before starting the measurement.

2. Check the positioning indicator.

Your monitor has a built-in advanced positioning sensor that is used as an aid in determining if the monitor is at the correct height. It has been designed to work with most people so that when your wrist is at the correct position relative to your heart,

the " \P " symbol will appear. When your wrist is higher or lower position relative to your heart, a locator bar will appear according to the wrist position. Due to difference in individual size and physique, this feature may not be helpful in all cases and you may wish to disable this feature. If you feel the position of the wrist according to positioning indicator's guidance does NOT match your heart level, disable this feature and follow your judgement. To disable it, refer to section 5.1.



	Monitor height	Positioning indicator	
	Wrist is too high	Any locator bar (-) above the " \overline{ullet} " symbol appears.	
•	Proper position	The " 🛡 " symbol appears.	
	Wrist is too low	Any locator bar (-) below the " $\overline{\bullet}$ " symbol appears.	

• After 5 seconds from pressing the [START/STOP] button, even if your monitor is not positioned properly, your monitor will start the measurement and the wrist cuff will start to inflate

3. Remain still and do not move or talk until the entire measurement process is completed. As the cuff inflates, your monitor automatically determines your

ideal inflation level. This monitor detects your blood pressure and pulse rate during inflation.

The " 🏶 " symbol flashes at every heartbeat.

• The " 🛞 " symbol appears if the wrist cuff is wrapped around the wrist correctly.

• When the " 🕥 " symbol is displayed, the wrist cuff is not applied correctly. Press the [START/STOP] button to turn your monitor off, then apply the cuff correctly.

After your monitor has detected your blood pressure and pulse rate, the cuff automatically deflates. Your blood pressure and pulse rate are displayed.


If either the systolic or the diastolic reading is high (refer to section 1.3), the " MA " symbol appears.




4. Press the [START/STOP] button to turn your monitor off.

• Your monitor will automatically turn off after 2 minutes. • Wait 2-3 minutes between measurements. The wait time allows the arteries to decompress and return to their premeasurement form. You may need to increase the wait time depending on your individual physiological characteristics.








4. Using Memory Function

Your monitor automatically stores up to 60 readings.

4.1 Viewing the Readings Stored in Memory

1. Press the 🕑 button.



The Memory number appears for one second before the pulse rate is displayed. The most recent reading set is numbered "1"

- Note • If the positioning indicator is enabled, the positioning
- indicator is displayed with the readings.
- If the reading is high (refer to section 1.3),
- the " I appears. --· If there are no readings stored in the - memory, the screen to the right is displayed. - -
- If the memory is full, the monitor will delete the oldest readings.
- **2.** Press the **button** repeatedly to scroll through the

previous readings stored in the memory.

4.2 Viewing the Average Reading Your monitor can calculate an average reading based on the latest 2 or 3 readings taken within 10 minutes span.

- **1.** When your monitor is off, press and hold the button for more than 2 seconds
 - If there are only 2 readings in the memory for that period, the average will be based on these 2 readings.

 If the average reading is high (refer to section 1.3), the " I > " symbol appears.



Your monitor calculates and displays weekly averages for the measurements taken in the morning for 4 weeks.

1. Press the 🖾 button once. The morning weekly average for the current week "THIS WEEK" appears on the display. If the morning weekly average is high (refer to section 1.3), the " 🚆 " symbol appears.

₩ AVG 84 58

메 AVG

149


88

FΓ

- **2.** Press the button repeatedly to display the average for the previous weeks. Your monitor displays "-1 WEEK" for the previous week's average, to "-3 WEEK" for the oldest set of averages.
- 3. Press the [START/STOP] button to turn off your monitor.


Morning Weekly Average Calculation

This is the average for the measurements taken during the morning (4:00 - 9:59) between Sunday and Saturday. Only the first 3 measurements taken in the morning between 4:00 - 9:59 will be used to calculate the morning average for each day.



5. Other Settings

- 5.1 Disabling/Enabling Positioning Indicator
- The positioning indicator is enabled as default. **1.** When your monitor is off, press and hold the B button



- Your monitor will automatically turn off after 3 seconds from the completion
- To enable the positioning indicator, follow the same steps as above.

Ø

5.2 Deleting All Readings Stored in Memory

- **1.** Press the \bigcirc button, then the " \bigcirc " symbol appears.
- **2.** While holding the (🖻 B button down, press and hold the [START/ STOP] button for more + than 2 seconds. STAR

Note • All readings will be deleted. You cannot partially delete the readings stored in the memory.

6. Error Messages and Troubleshooting

In case of any of the below problems occur during measurement, first check that no other electrical device is within 30cm. If the problem persists, please refer to the table below

Display/ Problem	Possible Cause	Solution	
E appears or the	The wrist cuff is not applied correctly.	Apply the wrist cuff correctly, then take another measurement. Refer to section 2.4.	
wrist cuff does not inflate.	Air is leaking from the wrist cuff.	Contact your OMRON retail outlet or distributor.	
E appears	The wrist cuff is overinflated exceeding 300 mmHg.	Do not touch the wrist cuff while taking a measurement.	
E H appears	You move or talk during a measurement. Vibrations disrupt a measurement.	Remain still and do not talk during a measurement.	
ES appears		Apply the wrist cuff correctly, then take another measurement. Refer to section 2.4.	
appears	The pulse rate is not detected correctly.	Remain still and sit correctly during a measurement.	
appears		If the " ② " symbol continues to appear, we recommend you to consult with your physician.	
E appears	Your wrist is moved up and down during a measurement.	Do not move your wrist, then take another measurement. Refer to chapter 3.	
E r appears	The monitor is malfunctioned.	Press the [START/ STOP] button again. If "Er" still appears, contact your OMRON retail outlet or distributor.	
heart symbol The heart symbol in the positioning indicator does not appear and a locator bar does not appear.	The positioning indicator is disabled.	Enable the positioning indicator. Refer to section 5.1.	
flashes	Batteries are low.	Replacing all batteries with 2 new alkaline batteries is recommended. Refer to section 2.1.	

Display/ Problem	Possible Cause	Solution	
appears or the monitor is turned off unexpectedly during a measurement	Batteries are depleted.	Immediately replace all batteries with 2 new alkaline batteries. Refer to section 2.1.	
No power. Nothing	Batteries are completely depleted.		
appears on the display of the monitor.	Battery polarities are not properly aligned.	Check the battery installation for proper placement. Refer to section 2.1.	
Readings appear too high or too low.	Blood pressure varies constantly. Many factors including stress, time of day, and/or how you apply the wrist cuff, may affect your blood pressure. Review sections 2.3 - 2.5 and chapter 3.		
Any other problems occur.	Press the [START/STOP] button to turn the monitor off, then press it again to take a measurement. If the problem continues, remove all batteries and wait for 30 seconds. Then re- install batteries. If the problem still persists, contact your OMRON retail outlet or distributor.		

7. Maintenance

7.1 Maintenance

To protect your monitor from damage, please follow the directions below

• Changes or modifications not approved by the manufacturer will void the user warranty.

ACaution

• DO NOT disassemble or attempt to repair this monitor or other components. This may cause an inaccurate reading.

7.2 Storage

Keep your monitor in the storage case when not in use.

· Store your monitor in a clean, safe location. Do not store your monitor:

• If your monitor is wet.

- In locations exposed to extreme temperatures, humidity, direct sunlight, dust or corrosive vapors such as bleach.
- In locations exposed to vibrations or shocks.

7.3 Cleaning

- Do not use any abrasive or volatile cleaners. • Use a soft dry cloth or a soft cloth moistened with neutral soap to clean your monitor and wrist cuff, and then wipe them with a dry
- Do not wash or immerse your monitor and wrist cuff in water. • Do not use gasoline, thinners or similar solvents to clean your monitor and wrist cuff.

7.4 Calibration and Service

• The accuracy of this blood pressure monitor has been carefully tested and is designed for a long service life.

• It is generally recommended to have the unit inspected every two years to ensure correct functioning and accuracy. Please consult your authorised OMRON dealer or the OMRON Customer Service at the address given on the packaging or attached literature.

circumference Memory Contents

electric shock

of the applied part

- Note
 - determination of diastolic blood pressure.

 - finger, and against oblique falling water drops which may cause issues during a normal operation.

9. Limited Warranty

Thank you for buying an OMRON product. This product is constructed of high quality materials and great care has been taken in its manufacturing. It is designed to give you every satisfaction, provided that it is properly operated and maintained as described in the instruction manual. This product is warranted by OMRON for a period of 3 years after the date of purchase. The proper construction, workmanship and materials of this product is warranted by OMRON. During this period of warranty OMRON will, without charge for labour or parts, repair or replace the defect product or any defective parts. The warranty does not cover any of the following: A. Transport costs and risks of transport.

- unauthorised persons.
- C. Periodic check-ups and maintenance.
- charged for).
- from misuse

www.omron-healthcare.com

extension or renewal of the warranty period. consumer by the retailer.

Model (Code) Display Cuff pressure range Blood pressure

8. Specification

Product category

Product description

- measurement range Pulse measurement
- range Accuracy
- Inflation Deflation Measurement method Operation mode
- IP classification Rating Power source

Battery life

Durable period (Servie life) Operating conditions

Storage / Transport conditions

Weight

Dimensions

Measurable wrist

Protection against

Applied part Maximum temperature

ns	
	Electronic Sphygmomanometers
	Automatic Wrist Blood Pressure Monitor
	RS4 (HEM-6181-E)
	LCD digital display
	0 to 299 mmHg
	SYS: 60 to 260 mmHg
	DIA: 40 to 215 mmHg
	40 to 180 beats / min.
	Pressure: ±3 mmHg
	Pulse: ±5% of display reading
	Automatic by electric pump
	Automatic rapid deflation
d	Oscillometric method
	Continuous operation
	IP 22
	DC3 V 3.0 W
	2 "AAA" alkaline batteries 1.5V
	Approximately 300 measurements (using
	new alkaline batteries)
се	5 years
	+10°C to +40°C / 15 to 90% RH (non-
	condensing) / 800 to 1060 hPa
	-20°C to +60°C / 10 to 90% RH (non- condensing)
	Approximately 86 g not including
	batteries
	Approximately 93 mm (w) × 62 mm (h) ×
	20 mm (l)
	(not including the wrist cuff)
	13.5 to 21.5 cm
	Stores up to 60 readings
	Monitor, storage case, 2 "AAA" alkaline batteries, instruction manual
	Internally powered ME equipment
	Type BF (wrist cuff)
e	Lower than +48°C

• These specifications are subject to change without notice. • In the clinical validation study, K5 was used on 85 subjects for

• This monitor is clinically investigated according to the requirements of ISO 81060-2:2013.

• IP classification is degrees of protection provided by enclosures in accordance with IEC 60529. This monitor is protected against solid foreign objects of 12.5 mm diameter and greater such as a

• This device has not been validated for use on pregnant patients.

B. Costs for repairs and / or defects resulting from repairs done by

D. Failure or wear of optional parts or other attachments other than the main device itself, unless explicitly warranted above. E. Costs arising due to non-acceptance of a claim (those will be

F. Damages of any kind including personal caused accidentally or

G. Calibration service is not included within the warranty.

Should warranty service be required please apply to the dealer whom the product was purchased from or an authorised OMRON distributor. For the address refer to the product packaging / literature or to your specialised retailer. If you have difficulties in finding OMRON customer services, contact us for information.

Repair or replacement under the warranty does not give rise to any

The warranty will be granted only if the complete product is returned together with the original invoice / cash ticket issued to the

10. Guidance and Manufacturer's Declaration **C**€0197

• This device fulfils the provisions of EC directive 93/42/EEC

(Medical Device Directive). . This blood pressure monitor is designed according to the European Standard EN1060, Non-invasive sphygmomanometers Part 1: General Requirements and Part 3: Supplementary requirements for electromechanical blood pressure measuring systems.

• This OMRON product is produced under the strict quality system of OMRON HEALTHCARE Co., Ltd., Japan. The Core component for OMRON blood pressure monitors, which is the Pressure Sensor, is produced in Japan

Symbols description		
*	Applied part - Type BF Degree of protection against electric shock (leakage current)	
IP XX	Ingress protection degree provided by IEC 60529	
CE	CE Marking	
SN	Serial number	
ł	Temperature limitation	
	Humidity limitation	
	Atmospheric pressure limitation	
Ĩ	Need for the user to consult this instruction manual.	
13.5-21.5cm	Indicates the correct positioning for the monitor on the wrist Measurable wrist circumference	
۹	Battery	
	Direct current	

Important information regarding Electro Magnetic Compatibility (EMC)


HEM-6181-E manufactured by OMRON HEALTHCARE Co., Ltd. conforms to EN60601-1-2:2015 Electro Magnetic Compatibility (EMC) standard. Further documentation in accordance with this EMC standard is available at www.omron-healthcare.com. Refer to the EMC information for HEM-6181-E on the

Correct Disposal of This Product (Waste Electrical & Electronic

⋤

This marking shown on the product or its literature, indicates that it 🚿 should not be disposed of, with other household wastes at the end of its working life.

To prevent possible harm to the environment or human health from uncontrolled waste disposal, please separate this product from other types of wastes and recycle it responsibly to promote the stainable reuse of material resources



Manufacturer	OMRON HEALTHCARE Co., Ltd. 53, Kunotsubo, Terado-cho, Muko, KYOTO, 617-0002 JAPAN
EU-representative	OMRON HEALTHCARE EUROPE B.V. Scorpius 33, 2132 LR Hoofddorp, THE NETHERLANDS www.omron-healthcare.com
Production facility	OMRON HEALTHCARE MANUFACTURING VIETNAM CO., LTD. No.28 VSIP II, Street 2, Vietnam-Singapore Industrial Park II, Binh Duong Industry-Services-Urban Complex, Hoa Phu Ward, Thu Dau Mot City, Binh Duong Province, Vietnam
	OMRON HEALTHCARE UK LTD. Opal Drive, Fox Milne, Milton Keynes, MK15 0DG, UK www.omron-healthcare.com
Subsidiaries	OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH Konrad-Zuse-Ring 28, 68163 Mannheim, GERMANY www.omron-healthcare.com
	OMRON SANTÉ FRANCE SAS 14, rue de Lisbonne, 93561 Rosny-sous-Bois Cedex, FRANCE www.omron-healthcare.com

Equipment)